

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
 Data File : VU056318.D
 Acq On : 14 Nov 2023 20:18
 Operator : MD/SY
 Sample : 05333-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YCG95

Quant Time: Nov 15 02:57:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 01:36:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	270601	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	277680	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	119346	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	56036	2.769	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.400%
7) Chloroethane-d5	1.907	69	50505	3.254	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	65.000%
11) 1,1-Dichloroethene-d2	2.563	65	30677	3.434	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.600%
20) 2-Butanone-d5	4.618	46	171274	42.438	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.880%
24) Chloroform-d	5.058	84	149825	3.827	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.600%
26) 1,2-Dichloroethane-d4	5.698	65	73965	4.089	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.800%
32) Benzene-d6	5.724	84	288495	3.487	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	69.800%#
36) 1,2-Dichloropropane-d6	6.685	67	92819	3.723	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.400%
41) Toluene-d8	7.894	98	243626	3.335	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	66.600%#
43) trans-1,3-Dichloroprop...	8.177	79	31996	3.346	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.000%
46) 2-Hexanone-d5	8.627	63	145373	38.683	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.360%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	70277	3.944	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	88015	3.992	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	79.800%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111323\
Data File : VU056318.D
Acq On : 14 Nov 2023 20:18
Operator : MD/SY
Sample : 05333-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
YCG95

Quant Time: Nov 15 02:57:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 01:36:14 2023
Response via : Initial Calibration

