

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111418\
 Data File : VU028127.D
 Acq On : 14 Nov 2018 17:48
 Operator : MD/SY
 Sample : J5948-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9C25

Quant Time: Nov 21 18:18:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Nov 15 04:25:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	149554	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	139883	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	63962	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	55875	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.68	69	45575	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.40%
11) 1,1-Dichloroethene-d2	2.28	63	78017	3.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	4.19	46	182931	50.82	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.64%
24) Chloroform-d	4.65	84	92952	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	5.31	65	52584	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.34	84	191101	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
36) 1,2-Dichloropropane-d6	6.33	67	69315	4.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.60%
41) Toluene-d8	7.57	98	171785	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloropropene-	7.85	79	22373	4.37	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.40%
46) 2-Hexanone-d5	8.31	63	144025	49.21	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.42%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	49093	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	59257	4.96	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.20%

Target Compounds					Ovalue
47) Tetrachloroethene	8.23	164	2848	0.301 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111418\
Data File : VU028127.D
Acq On : 14 Nov 2018 17:48
Operator : MD/SY
Sample : J5948-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9C25

Quant Time: Nov 21 18:18:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 15 04:25:02 2018
Response via : Initial Calibration

