

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
 Data File : VU056344.D
 Acq On : 15 Nov 2023 08:24
 Operator : MD/SY
 Sample : 05390-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0TJ7

Quant Time: Nov 16 01:27:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 03:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	271140	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	268867	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	114352	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	53926	2.659	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	53.200%
7) Chloroethane-d5	1.911	69	47909	3.081	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	61.600%#
11) 1,1-Dichloroethene-d2	2.566	65	27857	3.112	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.200%
20) 2-Butanone-d5	4.621	46	164219	40.609	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.220%
24) Chloroform-d	5.058	84	143509	3.658	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.200%
26) 1,2-Dichloroethane-d4	5.698	65	71277	3.932	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.600%
32) Benzene-d6	5.724	84	273965	3.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.400%#
36) 1,2-Dichloropropane-d6	6.685	67	87867	3.639	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	72.800%
41) Toluene-d8	7.894	98	223562	3.160	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.200%#
43) trans-1,3-Dichloroprop...	8.177	79	30200	3.262	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.200%
46) 2-Hexanone-d5	8.631	63	130047	35.739	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	71.480%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	68987	3.999	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	82438	3.902	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	78.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
Data File : VU056344.D
Acq On : 15 Nov 2023 08:24
Operator : MD/SY
Sample : 05390-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0TJ7

Quant Time: Nov 16 01:27:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 03:03:12 2023
Response via : Initial Calibration

