

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
 Data File : VU056326.D
 Acq On : 15 Nov 2023 01:09
 Operator : MD/SY
 Sample : VU1114WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK072

Quant Time: Nov 15 03:03:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 03:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	276569	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	267073	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	117593	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	91823	4.439	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.800%
7) Chloroethane-d5	1.907	69	78089	4.923	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.400%
11) 1,1-Dichloroethene-d2	2.563	65	41519	4.548	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	4.618	46	201351	48.814	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.620%
24) Chloroform-d	5.055	84	191768	4.793	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%
26) 1,2-Dichloroethane-d4	5.695	65	95756	5.179	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
32) Benzene-d6	5.721	84	365783	4.597	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	6.685	67	114560	4.777	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.600%
41) Toluene-d8	7.894	98	306352	4.360	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
43) trans-1,3-Dichloroprop...	8.177	79	41244	4.484	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.627	63	179344	49.618	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.240%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	82128	4.793	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	106116	4.885	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
Data File : VU056326.D
Acq On : 15 Nov 2023 01:09
Operator : MD/SY
Sample : VU1114WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK072

Quant Time: Nov 15 03:03:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 03:03:12 2023
Response via : Initial Calibration

