

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
 Data File : VU056336.D
 Acq On : 15 Nov 2023 05:11
 Operator : MD/SY
 Sample : 05391-01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4A06

Quant Time: Nov 16 01:25:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 03:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	285013	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	278187	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	122252	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	58622	2.750	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	55.000%
7) Chloroethane-d5	1.907	69	55256	3.381	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	67.600%
11) 1,1-Dichloroethene-d2	2.563	65	29115	3.095	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.800%
20) 2-Butanone-d5	4.618	46	155253	36.523	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.040%
24) Chloroform-d	5.058	84	152196	3.691	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.800%
26) 1,2-Dichloroethane-d4	5.695	65	74473	3.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
32) Benzene-d6	5.721	84	283595	3.422	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.400%#
36) 1,2-Dichloropropane-d6	6.685	67	90781	3.634	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	72.600%
41) Toluene-d8	7.894	98	235576	3.219	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.400%#
43) trans-1,3-Dichloroprop...	8.177	79	31288	3.266	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.400%
46) 2-Hexanone-d5	8.627	63	123965	32.927	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	65.860%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	65931	3.694	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	85246	3.775	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	75.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
Data File : VU056336.D
Acq On : 15 Nov 2023 05:11
Operator : MD/SY
Sample : 05391-01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4A06

Quant Time: Nov 16 01:25:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 03:03:12 2023
Response via : Initial Calibration

