

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
 Data File : VU056337.D
 Acq On : 15 Nov 2023 05:35
 Operator : MD/SY
 Sample : 05328-17
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A49Z7

Quant Time: Nov 16 01:25:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 03:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	278923	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	278751	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	122933	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	55919	2.680	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	53.600%
7) Chloroethane-d5	1.911	69	54509	3.408	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.200%
11) 1,1-Dichloroethene-d2	2.563	65	30744	3.339	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.800%
20) 2-Butanone-d5	4.618	46	148763	35.760	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	71.520%
24) Chloroform-d	5.055	84	149545	3.706	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.200%
26) 1,2-Dichloroethane-d4	5.695	65	69691	3.737	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.800%
32) Benzene-d6	5.721	84	273570	3.294	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	65.800%#
36) 1,2-Dichloropropane-d6	6.685	67	88143	3.521	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	70.400%
41) Toluene-d8	7.894	98	230961	3.149	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.000%#
43) trans-1,3-Dichloroprop...	8.177	79	31166	3.247	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.000%
46) 2-Hexanone-d5	8.627	63	120574	31.961	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	63.920%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	64365	3.599	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	83374	3.671	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	73.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
Data File : VU056337.D
Acq On : 15 Nov 2023 05:35
Operator : MD/SY
Sample : 05328-17
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A49Z7

Quant Time: Nov 16 01:25:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 03:03:12 2023
Response via : Initial Calibration

