

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
 Data File : VU056343.D
 Acq On : 15 Nov 2023 08:00
 Operator : MD/SY
 Sample : 05390-03
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0TJ5

Quant Time: Nov 16 01:26:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 15 03:03:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	279720	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	275304	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	119420	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	55118	2.634	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	52.600%
7) Chloroethane-d5	1.910	69	47006	2.930	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	58.600%#
11) 1,1-Dichloroethene-d2	2.563	65	27933	3.025	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.600%
20) 2-Butanone-d5	4.621	46	165231	39.606	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.220%
24) Chloroform-d	5.058	84	148809	3.677	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.600%
26) 1,2-Dichloroethane-d4	5.698	65	70218	3.755	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.200%
32) Benzene-d6	5.724	84	277530	3.384	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	67.600%#
36) 1,2-Dichloropropane-d6	6.685	67	91812	3.714	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.200%
41) Toluene-d8	7.894	98	234104	3.232	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.600%#
43) trans-1,3-Dichloroprop...	8.177	79	30334	3.199	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.000%
46) 2-Hexanone-d5	8.630	63	131282	35.235	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.480%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	67319	3.811	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	88668	4.019	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111423\
Data File : VU056343.D
Acq On : 15 Nov 2023 08:00
Operator : MD/SY
Sample : 05390-03
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0TJ5

Quant Time: Nov 16 01:26:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 15 03:03:12 2023
Response via : Initial Calibration

