

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111518\
 Data File : VU028177.D
 Acq On : 15 Nov 2018 17:26
 Operator : MD/SY
 Sample : J5946-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9C03

Quant Time: Nov 16 02:58:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 16 00:43:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	167980	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	162586	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	72476	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	52006	3.73	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.60%
7) Chloroethane-d5	1.68	69	46341	4.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.80%
11) 1,1-Dichloroethene-d2	2.28	63	79650	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	4.19	46	212473	52.56	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.12%
24) Chloroform-d	4.65	84	103685	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.31	65	59408	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	5.34	84	205236	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	6.33	67	76334	4.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.40%
41) Toluene-d8	7.57	98	174797	4.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.40%
43) trans-1,3-Dichloropropene-	7.85	79	22979	3.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.20%
46) 2-Hexanone-d5	8.31	63	159599	46.92	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.84%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	53783	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	11.86	152	61914	4.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.40%

Target Compounds					Qvalue
13) Acetone	2.33	43	3594	1.32 ug/L	95
47) Tetrachloroethene	8.23	164	6456	0.59 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111518\
Data File : VU028177.D
Acq On : 15 Nov 2018 17:26
Operator : MD/SY
Sample : J5946-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9C03

Quant Time: Nov 16 02:58:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 16 00:43:47 2018
Response via : Initial Calibration

