

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U111519\
 Data File : VU035720.D
 Acq On : 15 Nov 2019 12:36
 Operator : JC/SP
 Sample : K5856-09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GAQG6

Quant Time: Nov 16 02:10:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR110419WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 16 01:36:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	216545	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	240606	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	102403	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	112059	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.93	69	104855	5.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.00%
11) 1,1-Dichloroethene-d2	2.59	63	139707	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.71	46	312431	63.49	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	126.98%
24) Chloroform-d	5.11	84	203975	5.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.80%
26) 1,2-Dichloroethane-d4	5.75	65	114275	6.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.40%
32) Benzene-d6	5.77	84	421261	5.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.20%
36) 1,2-Dichloropropane-d6	6.74	67	143113	5.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.00%
41) Toluene-d8	7.93	98	342969	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	8.22	79	50620	5.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.40%
46) 2-Hexanone-d5	8.68	63	229559	59.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.78%
57) 1,1,2,2-Tetrachloroethane-	10.79	84	123880	6.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.60%#
64) 1,2-Dichlorobenzene-d4	12.22	152	126456	6.23	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.60%#

Target Compounds				Ovalue
13) Acetone	2.68	43	9397	2.648 ug/L 94
34) Trichloroethene	6.58	95	21496	0.992 ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111519\
Data File : VU035720.D
Acq On : 15 Nov 2019 12:36
Operator : JC/SP
Sample : K5856-09
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GAQG6

Quant Time: Nov 16 02:10:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR110419WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 16 01:36:54 2019
Response via : Initial Calibration

