

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111521\
 Data File : VU045792.D
 Acq On : 15 Nov 2021 15:03
 Operator : SY/MD
 Sample : VU1115WBL04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK032

Quant Time: Nov 16 03:10:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 16 03:04:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	97073	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	97990	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	42824	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	36117	4.696	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.000%
7) Chloroethane-d5	1.916	69	28907	5.163	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.200%
11) 1,1-Dichloroethene-d2	2.572	65	15919	4.935	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.600%
20) 2-Butanone-d5	4.649	46	102521	49.742	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.480%
24) Chloroform-d	5.070	84	62283	4.861	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	5.707	65	37307	4.962	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.200%
32) Benzene-d6	5.732	84	133660	4.953	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	6.694	67	42078	4.948	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.903	98	115270	4.722	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
43) trans-1,3-Dichloroprop...	8.182	79	16680	4.831	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.600%
46) 2-Hexanone-d5	8.636	63	74663	48.047	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.100%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	35727	4.972	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	41028	5.580	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.600%
Target Compounds						Qvalue
16) Methylene chloride	3.051	84	1667	0.188	ug/L	97
25) Chloroform	5.092	83	3259	0.249	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111521\
Data File : VU045792.D
Acq On : 15 Nov 2021 15:03
Operator : SY/MD
Sample : VU1115WBL04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK032

Quant Time: Nov 16 03:10:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 16 03:04:04 2021
Response via : Initial Calibration

