

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
 Data File : VU056367.D
 Acq On : 15 Nov 2023 20:04
 Operator : MD/SY
 Sample : 05372-07
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCDQ3

Quant Time: Nov 16 04:42:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 16 04:37:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	281579	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	281909	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	123481	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	113845	5.405	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 108.200%		
7) Chloroethane-d5	1.911	69	88199	5.462	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 109.200%		
11) 1,1-Dichloroethene-d2	2.563	65	50823	5.468	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 109.400%		
20) 2-Butanone-d5	4.621	46	253869	60.450	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 120.900%		
24) Chloroform-d	5.055	84	219140	5.379	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 107.600%		
26) 1,2-Dichloroethane-d4	5.695	65	111160	5.905	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 118.200%		
32) Benzene-d6	5.721	84	440415	5.244	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.800%		
36) 1,2-Dichloropropane-d6	6.685	67	135521	5.354	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 107.000%		
41) Toluene-d8	7.894	98	378883	5.108	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 102.200%		
43) trans-1,3-Dichloroprop...	8.177	79	49121	5.060	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 101.200%		
46) 2-Hexanone-d5	8.631	63	209023	54.786	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 109.580%		
56) 1,1,2,2-Tetrachloroeth...	10.753	84	96809	5.352	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 107.000%		
66) 1,2-Dichlorobenzene-d4	12.190	152	132235	5.797	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 116.000%		
Target Compounds						Qvalue
13) Acetone	2.637	43	24664	6.895	ug/L	96
34) Trichloroethene	6.544	95	13101	0.598	ug/L	96
47) Tetrachloroethene	8.553	164	4923	0.333	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
Data File : VU056367.D
Acq On : 15 Nov 2023 20:04
Operator : MD/SY
Sample : 05372-07
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCDQ3

Quant Time: Nov 16 04:42:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 16 04:37:29 2023
Response via : Initial Calibration

