

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
 Data File : VU056379.D
 Acq On : 16 Nov 2023 01:44
 Operator : MD/SY
 Sample : 05411-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QG8

Quant Time: Nov 16 06:14:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 16 05:41:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	268793	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	273866	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	118275	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	106412	5.293	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.800%
7) Chloroethane-d5	1.907	69	89051	5.777	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.600%
11) 1,1-Dichloroethene-d2	2.563	65	49347	5.562	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	111.200%
20) 2-Butanone-d5	4.618	46	251722	62.790	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.580%
24) Chloroform-d	5.055	84	220383	5.667	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.400%
26) 1,2-Dichloroethane-d4	5.695	65	111499	6.205	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	124.000%
32) Benzene-d6	5.721	84	432313	5.298	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%
36) 1,2-Dichloropropane-d6	6.685	67	131855	5.362	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.200%
41) Toluene-d8	7.894	98	370030	5.136	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.800%
43) trans-1,3-Dichloroprop...	8.177	79	48221	5.113	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	102.200%
46) 2-Hexanone-d5	8.627	63	205588	55.468	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.940%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	96869	5.513	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	128907	5.900	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
Data File : VU056379.D
Acq On : 16 Nov 2023 01:44
Operator : MD/SY
Sample : 05411-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QG8

Quant Time: Nov 16 06:14:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 16 05:41:06 2023
Response via : Initial Calibration

