

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
 Data File : VU056386.D
 Acq On : 16 Nov 2023 04:34
 Operator : MD/SY
 Sample : 05411-13
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QJ2

Quant Time: Nov 16 06:16:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 16 05:41:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	265341	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	264727	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	116706	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	92577	4.665	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.200%
7) Chloroethane-d5	1.911	69	81544	5.359	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.200%
11) 1,1-Dichloroethene-d2	2.564	65	47524	5.426	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	108.600%
20) 2-Butanone-d5	4.618	46	229470	57.984	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.960%
24) Chloroform-d	5.056	84	199190	5.189	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.800%
26) 1,2-Dichloroethane-d4	5.695	65	98519	5.554	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.000%
32) Benzene-d6	5.721	84	395474	5.014	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
36) 1,2-Dichloropropane-d6	6.686	67	122190	5.140	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.800%
41) Toluene-d8	7.895	98	338087	4.854	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
43) trans-1,3-Dichloroprop...	8.174	79	45415	4.981	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.600%
46) 2-Hexanone-d5	8.628	63	184979	51.631	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.260%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	91864	5.408	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	125222	5.808	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
Data File : VU056386.D
Acq On : 16 Nov 2023 04:34
Operator : MD/SY
Sample : 05411-13
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QJ2

Quant Time: Nov 16 06:16:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 16 05:41:06 2023
Response via : Initial Calibration

