

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
 Data File : VU056388.D
 Acq On : 16 Nov 2023 05:23
 Operator : MD/SY
 Sample : 05411-15
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QK0

Quant Time: Nov 16 06:17:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 16 05:41:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	280671	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	291986	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	124812	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	94113	4.483	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.600%
7) Chloroethane-d5	1.911	69	85873	5.335	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.600%
11) 1,1-Dichloroethene-d2	2.563	65	49196	5.310	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.200%
20) 2-Butanone-d5	4.621	46	225594	53.892	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.780%
24) Chloroform-d	5.055	84	206971	5.097	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	5.695	65	103343	5.508	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%
32) Benzene-d6	5.721	84	406821	4.676	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
36) 1,2-Dichloropropane-d6	6.685	67	124746	4.758	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.894	98	350573	4.564	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%
43) trans-1,3-Dichloroprop...	8.174	79	45999	4.574	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.627	63	194260	49.159	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.320%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	94178	5.027	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	127319	5.522	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111523\
Data File : VU056388.D
Acq On : 16 Nov 2023 05:23
Operator : MD/SY
Sample : 05411-15
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QK0

Quant Time: Nov 16 06:17:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 16 05:41:06 2023
Response via : Initial Calibration

