

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111618\
 Data File : VU028200.D
 Acq On : 16 Nov 2018 21:45
 Operator : MD/SY
 Sample : J5960-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AK6

Quant Time: Nov 17 03:33:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111418WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 16 23:59:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	158216	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	151354	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	68125	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	43500	3.31	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.20%
7) Chloroethane-d5	1.68	69	40442	3.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.60%
11) 1,1-Dichloroethene-d2	2.28	63	65811	2.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	53.60%#
20) 2-Butanone-d5	4.22	46	196841	51.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.38%
24) Chloroform-d	4.65	84	86843	4.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.20%
26) 1,2-Dichloroethane-d4	5.31	65	52762	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
32) Benzene-d6	5.34	84	170258	3.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.40%
36) 1,2-Dichloropropane-d6	6.33	67	66947	4.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.20%
41) Toluene-d8	7.57	98	140741	3.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.60%#
43) trans-1,3-Dichloropropene-	7.85	79	17046	3.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.60%
46) 2-Hexanone-d5	8.32	63	141781	44.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.54%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	48342	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	50973	4.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111618\
Data File : VU028200.D
Acq On : 16 Nov 2018 21:45
Operator : MD/SY
Sample : J5960-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AK6

Quant Time: Nov 17 03:33:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111418WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 16 23:59:22 2018
Response via : Initial Calibration

