

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111620\
 Data File : VU041239.D
 Acq On : 16 Nov 2020 10:33
 Operator : SY/MD
 Sample : I.BLK
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK21

Quant Time: Nov 16 10:51:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Nov 16 10:17:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	75386	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	82321	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	34872	5.00	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) Vinyl Chloride-d3	1.61	65	25644	4.05	ug/L	0.00
7) Chloroethane-d5	1.93	69	22535	4.63	ug/L	0.00
11) 1,1-Dichloroethene-d2	2.58	63	40380	3.25	ug/L	0.00
20) 2-Butanone-d5	4.66	46	107799	51.69	ug/L	0.00
24) Chloroform-d	5.08	84	53837	4.81	ug/L	0.00
26) 1,2-Dichloroethane-d4	5.72	65	35653	5.14	ug/L	0.00
32) Benzene-d6	5.74	84	96423	4.38	ug/L	0.00
36) 1,2-Dichloropropane-d6	6.70	67	33134	4.62	ug/L	0.00
41) Toluene-d8	7.91	98	86283	4.33	ug/L	0.00
43) trans-1,3-Dichloropropene-	8.19	79	15057	4.85	ug/L	0.00
46) 2-Hexanone-d5	8.64	63	77487	46.32	ug/L	0.00
57) 1,1,2,2-Tetrachloroethane-	10.75	84	30841	4.63	ug/L	0.00
64) 1,2-Dichlorobenzene-d4	12.19	152	34614	5.66	ug/L	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111620\
Data File : VU041239.D
Acq On : 16 Nov 2020 10:33
Operator : SY/MD
Sample : I.BLK
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK21

Quant Time: Nov 16 10:51:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Nov 16 10:17:27 2020
Response via : Initial Calibration

