

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
 Data File : VU045801.D
 Acq On : 16 Nov 2021 11:01
 Operator : SY/MD
 Sample : VU1116WBL07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK035

Quant Time: Nov 16 17:08:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 16 03:04:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	92213	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	96106	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	41778	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	30714	4.204	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.000%
7) Chloroethane-d5	1.919	69	25810	4.853	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.575	65	14128	4.611	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	4.655	46	114640	58.553	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.100%
24) Chloroform-d	5.067	84	55705	4.577	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.706	65	36652	5.132	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.600%
32) Benzene-d6	5.732	84	120379	4.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	6.694	67	39815	4.773	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.902	98	101346	4.233	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.600%
43) trans-1,3-Dichloroprop...	8.182	79	15723	4.643	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.800%
46) 2-Hexanone-d5	8.639	63	84523	55.459	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.920%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	38673	5.488	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	39637	5.526	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
Data File : VU045801.D
Acq On : 16 Nov 2021 11:01
Operator : SY/MD
Sample : VU1116WBL07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK035

Quant Time: Nov 16 17:08:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 16 03:04:04 2021
Response via : Initial Calibration

