

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
 Data File : VU045808.D
 Acq On : 16 Nov 2021 15:39
 Operator : SY/MD
 Sample : M4686-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG1Z1

Quant Time: Nov 17 04:51:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 17 04:49:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	87078	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	90509	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	41533	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	26492	3.840	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.800%
7) Chloroethane-d5	1.919	69	23090	4.597	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.572	65	12228	4.226	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.600%
20) 2-Butanone-d5	4.652	46	107843	58.330	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	116.660%
24) Chloroform-d	5.067	84	52630	4.579	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.707	65	34108	5.057	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.732	84	110478	4.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.694	67	36667	4.668	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.899	98	93394	4.142	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
43) trans-1,3-Dichloroprop...	8.182	79	14346	4.499	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.639	63	76156	53.059	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.120%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34020	5.126	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	37470	5.254	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.000%
Target Compounds						
					Qvalue	
3) Chloromethane	1.520	50	3012	0.401	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
Data File : VU045808.D
Acq On : 16 Nov 2021 15:39
Operator : SY/MD
Sample : M4686-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG1Z1

Quant Time: Nov 17 04:51:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 17 04:49:52 2021
Response via : Initial Calibration

