

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
 Data File : VU045823.D
 Acq On : 16 Nov 2021 21:38
 Operator : SY/MD
 Sample : M4686-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG214

Quant Time: Nov 17 06:26:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 17 04:49:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	94794	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	99854	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	46804	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	24112	3.210	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.200%
7) Chloroethane-d5	1.919	69	21327	3.901	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.575	65	11175	3.548	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.000%
20) 2-Butanone-d5	4.645	46	109882	54.595	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.180%
24) Chloroform-d	5.067	84	46139	3.687	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.800%
26) 1,2-Dichloroethane-d4	5.707	65	32070	4.368	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
32) Benzene-d6	5.732	84	104702	3.808	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.200%
36) 1,2-Dichloropropane-d6	6.694	67	35363	4.080	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.600%
41) Toluene-d8	7.903	98	88452	3.556	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.200%
43) trans-1,3-Dichloroprop...	8.182	79	13768	3.913	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.636	63	77631	49.025	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.040%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	33677	4.600	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	37250	4.635	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.388	85	5698	0.738	ug/L	95
9) Trichlorofluoromethane	2.144	101	609	0.057	ug/L	# 82
17) Methyl tert-butyl Ether	3.372	73	10509	0.688	ug/L	98
22) cis-1,2-Dichloroethene	4.671	96	5863	0.860	ug/L	96
25) Chloroform	5.096	83	8439	0.660	ug/L	96
34) Trichloroethene	6.546	95	129590	18.046	ug/L	98
47) Tetrachloroethene	8.555	164	122613	24.390	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
Data File : VU045823.D
Acq On : 16 Nov 2021 21:38
Operator : SY/MD
Sample : M4686-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG214

Quant Time: Nov 17 06:26:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 17 04:49:52 2021
Response via : Initial Calibration

