

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
 Data File : VU045810.D
 Acq On : 16 Nov 2021 16:26
 Operator : SY/MD
 Sample : M4686-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG1Z3

Quant Time: Nov 17 04:51:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 17 04:49:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	90420	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	93965	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	43745	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	26692	3.726	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.600%
7) Chloroethane-d5	1.919	69	22945	4.400	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	2.571	65	11890	3.957	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	4.645	46	108415	56.472	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.940%
24) Chloroform-d	5.067	84	50463	4.228	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
26) 1,2-Dichloroethane-d4	5.706	65	33264	4.750	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
32) Benzene-d6	5.732	84	107776	4.165	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	6.694	67	37009	4.538	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.800%
41) Toluene-d8	7.899	98	93288	3.985	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.179	79	14143	4.272	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.400%
46) 2-Hexanone-d5	8.636	63	77797	52.209	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.420%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34782	5.048	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	36294	4.832	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%
Target Compounds						Qvalue
9) Trichlorofluoromethane	2.144	101	7221	0.714	ug/L	98
12) 1,1-Dichloroethene	2.581	96	763	0.136	ug/L	# 1
17) Methyl tert-butyl Ether	3.372	73	3199	0.220	ug/L	98
22) cis-1,2-Dichloroethene	4.678	96	5935	0.912	ug/L	96
25) Chloroform	5.092	83	8448	0.693	ug/L	94
31) Carbon tetrachloride	5.536	117	2717	0.316	ug/L	94
34) Trichloroethene	6.546	95	265248	39.251	ug/L	99
47) Tetrachloroethene	8.555	164	47144	9.965	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111621\
Data File : VU045810.D
Acq On : 16 Nov 2021 16:26
Operator : SY/MD
Sample : M4686-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG1Z3

Quant Time: Nov 17 04:51:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 17 04:49:52 2021
Response via : Initial Calibration









#9
Trichlorofluoromethane
Concen: 0.714 ug/L
RT: 2.144 min Scan# 21
Delta R.T. 0.000 min
Lab File: VU045810.D
Acq: 16 Nov 2021 16:26




Instrument :
MSVOA_U
ClientSampleId :
BG1Z3


Tgt Ion:101 Resp: 7221
Ion Ratio Lower Upper
101 100
103 63.0 51.5 77.3



#12
1,1-Dichloroethene
Concen: 0.136 ug/L
RT: 2.581 min Scan# 386
Delta R.T. -0.003 min
Lab File: VU045810.D
Acq: 16 Nov 2021 16:26

Tgt Ion: 96 Resp: 763
Ion Ratio Lower Upper
96 100
61 504.6 124.7 231.5#
63 3613.2 105.2 195.4#









#17
Methyl tert-butyl Ether
Concen: 0.220 ug/L
RT: 3.372 min Scan# 61
Delta R.T. 0.003 min
Lab File: VU045810.D
Acq: 16 Nov 2021 16:26




Instrument :
MSVOA_U
ClientSampleId :
BG1Z3


Tgt Ion: 73 Resp: 3199
Ion Ratio Lower Upper
73 100
43 17.6 14.8 22.2
57 22.5 17.4 26.0



#22
cis-1,2-Dichloroethene
Concen: 0.912 ug/L
RT: 4.678 min Scan# 1038
Delta R.T. 0.003 min
Lab File: VU045810.D
Acq: 16 Nov 2021 16:26

Tgt Ion: 96 Resp: 5935
Ion Ratio Lower Upper
96 100
61 120.5 88.5 164.3
98 65.6 44.4 82.5





#25

Chloroform

Concen: 0.693 ug/L

RT: 5.092 min Scan# 1167

Delta R.T. 0.000 min

Lab File: VU045810.D

Acq: 16 Nov 2021 16:26

Instrument :

MSVOA_U

ClientSampleId :





BG1Z3

Tgt Ion: 83 Resp: 8448

Ion Ratio Lower Upper

83 100

85 62.0 46.9 87.1



#31

Carbon tetrachloride

Concen: 0.316 ug/L

RT: 5.536 min Scan# 1305

Delta R.T. 0.003 min

Lab File: VU045810.D




Acq: 16 Nov 2021 16:26


Tgt Ion: 117 Resp: 2717

Ion Ratio Lower Upper

117 100

119 102.5 77.5 116.3









#34
Trichloroethene
Concen: 39.251 ug/L
RT: 6.546 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VU045810.D
Acq: 16 Nov 2021 16:26

Instrument :
MSVOA_U
ClientSampleId :
BG1Z3

Tgt Ion: 95 Resp: 265248

Ion	Ratio	Lower	Upper
95	100		
97	64.8	46.5	86.5
132	93.3	65.1	120.9
130	95.8	67.3	124.9



#47
Tetrachloroethene
Concen: 9.965 ug/L
RT: 8.555 min Scan# 2244
Delta R.T. 0.000 min
Lab File: VU045810.D
Acq: 16 Nov 2021 16:26

Tgt Ion:164 Resp: 47144

Ion	Ratio	Lower	Upper
164	100		
129	98.3	71.8	133.4
131	91.9	68.5	127.3
166	124.1	94.1	174.8

