

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111623\
 Data File : VU056426.D
 Acq On : 17 Nov 2023 05:20
 Operator : MD/SY
 Sample : 05341-09
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QK4

Quant Time: Nov 17 06:13:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 04:03:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	257056	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	253026	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	111707	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	80586	4.191	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.800%
7) Chloroethane-d5	1.908	69	70907	4.810	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.200%
11) 1,1-Dichloroethene-d2	2.563	65	39290	4.630	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.600%
20) 2-Butanone-d5	4.621	46	219581	57.274	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.540%
24) Chloroform-d	5.055	84	187835	5.051	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
26) 1,2-Dichloroethane-d4	5.695	65	94526	5.501	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.000%
32) Benzene-d6	5.721	84	361181	4.791	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%
36) 1,2-Dichloropropane-d6	6.685	67	115302	5.075	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.400%
41) Toluene-d8	7.894	98	304970	4.581	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.600%
43) trans-1,3-Dichloroprop...	8.174	79	39248	4.504	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.631	63	179196	52.330	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.660%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	83843	5.164	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	109947	5.328	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.600%
Target Compounds						
42) Toluene	7.965	91	21875	0.293	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111623\
Data File : VU056426.D
Acq On : 17 Nov 2023 05:20
Operator : MD/SY
Sample : 05341-09
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QK4

Quant Time: Nov 17 06:13:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 04:03:21 2023
Response via : Initial Calibration

