

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111623\
 Data File : VU056423.D
 Acq On : 17 Nov 2023 03:12
 Operator : MD/SY
 Sample : 05341-05
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QJ5

Quant Time: Nov 17 04:32:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 04:03:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	246205	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	244340	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	101440	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	76181	4.137	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.800%
7) Chloroethane-d5	1.908	69	68141	4.826	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.560	65	36520	4.494	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.800%
20) 2-Butanone-d5	4.618	46	229741	62.565	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.140%
24) Chloroform-d	5.055	84	191401	5.373	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.400%
26) 1,2-Dichloroethane-d4	5.695	65	98224	5.968	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.400%
32) Benzene-d6	5.721	84	371167	5.099	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
36) 1,2-Dichloropropane-d6	6.682	67	113316	5.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.200%
41) Toluene-d8	7.894	98	295802	4.601	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%
43) trans-1,3-Dichloroprop...	8.174	79	38749	4.605	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.628	63	174912	52.894	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.780%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	84378	5.382	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	107928	5.759	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111623\
Data File : VU056423.D
Acq On : 17 Nov 2023 03:12
Operator : MD/SY
Sample : 05341-05
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QJ5

Quant Time: Nov 17 04:32:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 04:03:21 2023
Response via : Initial Calibration

