

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111624\
 Data File : VU061735.D
 Acq On : 16 Nov 2024 10:50
 Operator : MD/SY
 Sample : P4800-11
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCPB2

Quant Time: Nov 18 00:18:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 18 00:17:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	179652	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	169586	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	77209	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	41141	3.544	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.800%
7) Chloroethane-d5	1.907	69	43159	4.355	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.200%
11) 1,1-Dichloroethene-d2	2.560	65	18949	3.638	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.800%
20) 2-Butanone-d5	4.621	46	106356	43.169	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.340%
24) Chloroform-d	5.052	84	93656	4.059	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
26) 1,2-Dichloroethane-d4	5.692	65	48935	4.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
32) Benzene-d6	5.717	84	190559	4.062	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
36) 1,2-Dichloropropane-d6	6.682	67	58745	4.245	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.000%
41) Toluene-d8	7.891	98	170359	3.896	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.000%
43) trans-1,3-Dichloroprop...	8.174	79	19352	3.797	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.000%
46) 2-Hexanone-d5	8.627	63	88201	43.792	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.580%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	47634	4.367	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	12.187	152	60632	4.409	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.200%
Target Compounds						
5) Vinyl chloride	1.599	62	7121	0.533	ug/L	89
10) 1,1,2-Trichloro-1,2,2-...	2.566	101	3122	0.269	ug/L	97
13) Acetone	2.634	43	22597	14.558	ug/L	98
22) cis-1,2-Dichloroethene	4.656	96	8500	0.634	ug/L	98
33) Benzene	5.766	78	19309	0.376	ug/L	100
34) Trichloroethene	6.534	95	52534	3.791	ug/L	95
47) Tetrachloroethene	8.544	164	60483	5.808	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111624\
Data File : VU061735.D
Acq On : 16 Nov 2024 10:50
Operator : MD/SY
Sample : P4800-11
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCPB2

Quant Time: Nov 18 00:18:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 18 00:17:37 2024
Response via : Initial Calibration

