

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111624\
 Data File : VU061756.D
 Acq On : 16 Nov 2024 20:41
 Operator : MD/SY
 Sample : VU1116WBL02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK088

Quant Time: Nov 18 04:18:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 18 04:15:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	170243	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	163240	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	69749	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	44503	4.046	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.000%
7) Chloroethane-d5	1.911	69	41471	4.416	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.400%
11) 1,1-Dichloroethene-d2	2.560	65	19726	3.997	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.000%
20) 2-Butanone-d5	4.631	46	124155	53.178	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.360%
24) Chloroform-d	5.052	84	100408	4.592	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
26) 1,2-Dichloroethane-d4	5.695	65	53526	5.101	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%
32) Benzene-d6	5.718	84	204505	4.528	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
36) 1,2-Dichloropropane-d6	6.682	67	65025	4.882	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.600%
41) Toluene-d8	7.891	98	173157	4.114	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
43) trans-1,3-Dichloroprop...	8.174	79	19618	3.999	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.000%
46) 2-Hexanone-d5	8.631	63	103211	53.237	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.480%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	52479	4.998	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	61414	4.944	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111624\
Data File : VU061756.D
Acq On : 16 Nov 2024 20:41
Operator : MD/SY
Sample : VU1116WBL02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK088

Quant Time: Nov 18 04:18:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 18 04:15:29 2024
Response via : Initial Calibration

