

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\U111718\
 Data File : VU028240.D
 Acq On : 17 Nov 2018 21:28
 Operator : MD/SY
 Sample : J5986-02MS
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0FG7MS

Quant Time: Nov 19 03:59:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Nov 19 03:08:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	155888	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	147967	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	68646	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	54634	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	1.68	69	45718	4.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.20%
11) 1,1-Dichloroethene-d2	2.28	63	110497	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.60%
20) 2-Butanone-d5	4.19	46	199415	50.80	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.60%
24) Chloroform-d	4.65	84	93859	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	5.31	65	58465	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.34	84	198512	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.33	67	73592	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.20%
41) Toluene-d8	7.57	98	171689	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	7.85	79	22734	4.60	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.00%
46) 2-Hexanone-d5	8.31	63	146996	49.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.22%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	53812	4.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	60389	4.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.60%

Target Compounds

						Qvalue
5) Vinyl chloride	1.40	62	154760	9.78	ug/L	96
12) 1,1-Dichloroethene	2.29	96	50328	5.24	ug/L	94
17) Methyl tert-butyl Ether	3.01	73	536150	18.60	ug/L	99
18) trans-1,2-Dichloroethene	2.98	96	36472	3.57	ug/L	96
19) 1,1-Dichloroethane	3.45	63	23911	1.01	ug/L	95
22) cis-1,2-Dichloroethene	4.23	96	1002163	82.94	ug/L	97
33) Benzene	5.39	78	256599	5.26	ug/L	100
34) Trichloroethene	6.19	95	1103232	92.30	ug/L	99
42) Toluene	7.64	91	240652	4.86	ug/L	98
47) Tetrachloroethene	8.23	164	10237	1.23	ug/L	99
51) Chlorobenzene	9.12	112	138852	4.87	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU111718\
Data File : VU028240.D
Acq On : 17 Nov 2018 21:28
Operator : MD/SY
Sample : J5986-02MS
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0FG7MS

Quant Time: Nov 19 03:59:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Nov 19 03:08:19 2018
Response via : Initial Calibration

