

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111720\
 Data File : VU041270.D
 Acq On : 17 Nov 2020 10:53
 Operator : SY/MD
 Sample : VU1117WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK23

Quant Time: Nov 17 22:41:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 17 22:32:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	70810	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	73884	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	31972	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	24121	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.92	69	20793	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.40%
11) 1,1-Dichloroethene-d2	2.58	63	38894	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	4.66	46	102041	47.51	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.02%
24) Chloroform-d	5.08	84	55388	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.72	65	35335	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.74	84	101173	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	6.70	67	35007	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.91	98	86088	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloropropene-	8.19	79	13540	4.58	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.60%
46) 2-Hexanone-d5	8.64	63	73823	46.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.56%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	30425	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	32991	5.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.80%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	954	0.165 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111720\
Data File : VU041270.D
Acq On : 17 Nov 2020 10:53
Operator : SY/MD
Sample : VU1117WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK23

Quant Time: Nov 17 22:41:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 17 22:32:58 2020
Response via : Initial Calibration

