


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111720\
 Data File : VU041268.D
 Acq On : 17 Nov 2020 09:05
 Operator : SY/MD
 Sample : BFB13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB13

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Nov 18 00:06:52 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	25928	PASS
75	95	30	80	56.4	69848	PASS
95	95	100	100	100.0	123834	PASS
96	95	5	9	6.4	7898	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	74.5	92242	PASS
175	174	5	9	7.5	6895	PASS
176	174	95	101	95.5	88053	PASS
177	176	5	9	6.6	5777	PASS

m/z	Abundance
36.00	1357.0
37.10	7169.0
38.10	6440.0
39.10	2608.0
40.10	239.0
44.00	997.0
45.10	1121.0
46.10	114.0
47.00	1735.0
48.00	834.0
49.10	6037.0
50.10	25808.0
51.10	8738.0
52.10	384.0
55.00	392.0
56.10	2319.0
57.10	3816.0
58.00	125.0
60.00	1361.0
61.10	7183.0
62.10	7244.0
63.10	6449.0
64.00	943.0
65.10	1159.0
66.00	151.0
67.00	389.0
68.10	13715.0
69.10	14269.0
70.10	1176.0
72.00	766.0
73.10	6224.0
74.10	23904.0
75.10	69272.0
76.10	6028.0
77.00	786.0
78.00	441.0
79.00	5445.0
80.00	1593.0
81.00	5327.0
82.00	1327.0
83.10	144.0
86.10	130.0
87.10	4736.0
88.00	4694.0
91.00	507.0
92.10	4180.0
93.10	5735.0
94.10	15816.0
95.10	122344.0
96.10	7829.0
97.10	228.0
103.90	611.0
105.00	188.0
106.00	631.0
107.00	103.0
111.00	158.0
113.00	141.0
114.90	146.0
115.90	523.0
117.00	991.0
117.90	539.0
119.00	979.0
128.10	491.0
129.10	207.0
130.00	499.0
131.00	149.0
135.00	184.0
137.00	254.0
141.00	1382.0
142.00	168.0
143.10	2136.0
144.00	151.0
145.00	610.0
145.90	216.0
148.10	322.0
149.10	117.0
149.90	143.0
153.00	107.0
155.00	322.0
157.10	228.0
159.00	162.0
171.10	322.0
172.10	1014.0
174.10	89272.0
175.10	6629.0
176.10	84888.0
177.10	5646.0
178.00	194.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13

m/z	Abundance
36.00	1369.0
37.00	7323.0
38.10	6450.0
39.10	2672.0
40.10	210.0
44.10	1077.0
45.00	1180.0
46.10	129.0
47.00	1643.0
48.00	922.0
49.10	6251.0
50.10	26688.0
51.10	8959.0
52.00	378.0
53.00	139.0
55.00	415.0
56.10	2407.0
57.10	3851.0
58.00	143.0
60.00	1375.0
61.10	7394.0
62.10	7697.0
63.10	6340.0
64.00	915.0
65.10	1162.0
66.10	142.0
67.00	373.0
68.10	14050.0
69.10	15105.0
70.10	1186.0
72.00	866.0
73.10	6510.0
74.10	24792.0
75.10	72368.0
76.10	6146.0
77.00	804.0
78.00	476.0
79.00	5446.0
80.00	1654.0
81.00	5569.0
82.00	1302.0
83.10	192.0
86.10	139.0
87.10	4692.0
88.00	4773.0
91.00	581.0
92.10	4449.0
93.10	6137.0
94.10	16400.0
95.10	128112.0
96.10	8156.0
97.10	220.0
103.90	716.0
104.90	215.0
106.00	685.0
109.90	106.0
110.90	142.0
113.00	115.0
114.90	135.0
115.90	530.0
117.00	1005.0
117.90	596.0
119.10	943.0
128.10	509.0
129.10	251.0
130.00	599.0
131.00	151.0
135.00	206.0
137.10	269.0
141.00	1461.0
142.00	184.0
143.00	2119.0
144.00	173.0
145.00	568.0
145.90	203.0
147.00	102.0
148.10	337.0
149.10	124.0
149.90	191.0
153.10	107.0
155.00	304.0
156.00	117.0
157.10	242.0
159.00	197.0
161.00	114.0
171.10	265.0
172.10	1064.0
174.10	95536.0
175.10	7093.0
176.10	91360.0
177.10	6056.0
178.00	219.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13

m/z	Abundance
36.00	1235.0
37.00	6842.0
38.00	5965.0
39.10	2492.0
40.00	195.0
44.10	1019.0
45.00	1151.0
46.00	125.0
47.10	1369.0
48.00	924.0
49.10	5781.0
50.10	25288.0
51.10	8160.0
52.10	333.0
53.00	159.0
55.00	373.0
56.00	2262.0
57.10	3551.0
58.00	168.0
60.00	1244.0
61.10	6957.0
62.10	7273.0
63.10	5645.0
64.00	796.0
65.10	1016.0
67.00	340.0
68.10	12966.0
69.10	14038.0
70.00	1141.0
72.00	903.0
73.10	6157.0
74.10	23136.0
75.10	67904.0
76.10	5732.0
77.00	707.0
78.00	465.0
79.00	5033.0
80.00	1547.0
81.00	5274.0
82.00	1129.0
83.10	191.0
86.00	153.0
87.10	4269.0
88.00	4403.0
91.00	598.0
92.10	4158.0
93.10	5933.0
94.10	15520.0
95.10	121048.0
96.10	7711.0
97.10	180.0
103.90	705.0
104.90	209.0
106.00	703.0
110.90	127.0
114.90	116.0
116.00	519.0
117.00	947.0
118.00	616.0
119.10	819.0
128.00	488.0
129.10	232.0
130.00	624.0
131.00	191.0
135.00	209.0
137.10	285.0
140.10	106.0
141.00	1388.0
142.00	185.0
143.00	2037.0
144.00	187.0
145.00	495.0
145.90	181.0
148.00	316.0
150.00	208.0
155.00	273.0
156.00	114.0
157.10	216.0
159.00	201.0
161.00	142.0
171.10	216.0
172.10	1070.0
174.10	91920.0
175.10	6964.0
176.10	87912.0
177.10	5631.0
178.00	192.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13

m/z	Abundance
44.00	151.0

Instrument :
MSVOA_U
ClientSampleId :
BFB13