

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111721\
 Data File : VU045846.D
 Acq On : 17 Nov 2021 17:40
 Operator : SY/MD
 Sample : VIBLK016
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK016

Quant Time: Nov 18 02:22:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 18 02:11:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	92637	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	98065	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	43930	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	27435	3.738	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.800%
7) Chloroethane-d5	1.916	69	23881	4.470	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.400%
11) 1,1-Dichloroethene-d2	2.571	65	12526	4.069	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.400%
20) 2-Butanone-d5	4.642	46	98171	49.912	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.820%
24) Chloroform-d	5.070	84	53090	4.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	5.706	65	33215	4.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.732	84	113590	4.206	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.694	67	36901	4.336	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.902	98	93376	3.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.400%
43) trans-1,3-Dichloroprop...	8.179	79	13125	3.799	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.000%
46) 2-Hexanone-d5	8.636	63	63504	40.835	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.660%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	32681	4.545	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	37214	4.934	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.600%
Target Compounds						
					Qvalue	
3) Chloromethane	1.520	50	1469	0.184	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111721\
Data File : VU045846.D
Acq On : 17 Nov 2021 17:40
Operator : SY/MD
Sample : VIBLK016
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK016

Quant Time: Nov 18 02:22:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 18 02:11:55 2021
Response via : Initial Calibration

