

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111722\
 Data File : VU051886.D
 Acq On : 17 Nov 2022 11:06
 Operator : JC/MD
 Sample : VU1117MBL07
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK019

Quant Time: Nov 18 00:54:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM111522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 18 00:53:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	479072	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	474901	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	182635	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	106650	45.346	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	90.700%	
7) Chloroethane-d5	1.906	69	114265	48.491	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.980%	
11) 1,1-Dichloroethene-d2	2.565	63	130411	30.921	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.840%	
21) 2-Butanone-d5	4.616	46	269548	100.788	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	100.790%	
24) Chloroform-d	5.060	84	278100	47.170	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.340%	
26) 1,2-Dichloroethane-d4	5.697	65	175855	48.726	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.460%	
32) Benzene-d6	5.726	84	488129	49.296	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.600%	
36) 1,2-Dichloropropane-d6	6.687	67	184431	50.039	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	100.080%	
41) Toluene-d8	7.896	98	371096	45.659	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.320%	
43) trans-1,3-Dichloroprop...	8.176	79	68571	46.708	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.420%	
47) 2-Hexanone-d5	8.629	63	186272	99.441	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	99.440%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	346920	48.362	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	96.720%	
66) 1,2-Dichlorobenzene-d4	12.192	152	178852	55.647	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	111.300%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111722\
Data File : VU051886.D
Acq On : 17 Nov 2022 11:06
Operator : JC/MD
Sample : VU1117MBL07
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK019

Quant Time: Nov 18 00:54:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM111522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 18 00:53:49 2022
Response via : Initial Calibration

