

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111722\
 Data File : VU051888.D
 Acq On : 17 Nov 2022 11:53
 Operator : JC/MD
 Sample : VU1117WBL04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK021

Quant Time: Nov 18 00:54:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM111522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 18 00:53:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	435439	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	441655	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	170358	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	94527	44.219	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	88.440%	
7) Chloroethane-d5	1.903	69	103352	48.255	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.500%	
11) 1,1-Dichloroethene-d2	2.562	63	115129	30.033	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	60.060%	
21) 2-Butanone-d5	4.613	46	277967	114.351	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	114.350%	
24) Chloroform-d	5.060	84	261448	48.789	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.580%	
26) 1,2-Dichloroethane-d4	5.700	65	169008	51.522	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.040%	
32) Benzene-d6	5.726	84	456418	49.564	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.120%	
36) 1,2-Dichloropropane-d6	6.687	67	179204	52.280	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.560%	
41) Toluene-d8	7.896	98	339698	44.942	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.880%	
43) trans-1,3-Dichloroprop...	8.176	79	65997	48.339	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.680%	
47) 2-Hexanone-d5	8.629	63	182507	104.765	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	104.760%	
56) 1,1,2,2-Tetrachloroeth...	10.754	84	337307	50.561	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.120%	
66) 1,2-Dichlorobenzene-d4	12.192	152	166789	55.633	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	111.260%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111722\
Data File : VU051888.D
Acq On : 17 Nov 2022 11:53
Operator : JC/MD
Sample : VU1117WBL04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK021

Quant Time: Nov 18 00:54:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM111522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 18 00:53:49 2022
Response via : Initial Calibration

