

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111723\
 Data File : VU056436.D
 Acq On : 17 Nov 2023 12:48
 Operator : MD/SY
 Sample : 05389-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YCGE1

Quant Time: Nov 17 23:20:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 23:14:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	262332	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	252327	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	114631	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	70372	3.586	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.800%	
7) Chloroethane-d5	1.907	69	71960	4.783	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.600%	
11) 1,1-Dichloroethene-d2	2.563	65	34198	3.949	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.000%	
20) 2-Butanone-d5	4.615	46	202646	51.794	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	103.580%	
24) Chloroform-d	5.055	84	178326	4.699	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
26) 1,2-Dichloroethane-d4	5.698	65	91437	5.214	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.200%	
32) Benzene-d6	5.721	84	345068	4.590	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.800%	
36) 1,2-Dichloropropane-d6	6.685	67	112651	4.972	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.400%	
41) Toluene-d8	7.894	98	277285	4.177	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	8.177	79	36534	4.204	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.000%	
46) 2-Hexanone-d5	8.627	63	168769	49.421	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.840%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	82880	5.119	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	103883	4.906	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.200%	
Target Compounds						
13) Acetone	2.624	43	20891	6.269	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111723\
 Data File : VU056436.D
 Acq On : 17 Nov 2023 12:48
 Operator : MD/SY
 Sample : 05389-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 YCGE1

Quant Time: Nov 17 23:20:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 23:14:39 2023
 Response via : Initial Calibration

