

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111723\
 Data File : VU056449.D
 Acq On : 17 Nov 2023 18:07
 Operator : MD/SY
 Sample : 05441-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QF9

Quant Time: Nov 17 23:23:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 17 23:14:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	275221	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	281358	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	116093	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	66602	3.235	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.800%
7) Chloroethane-d5	1.907	69	69348	4.394	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.800%
11) 1,1-Dichloroethene-d2	2.563	65	33161	3.650	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.000%
20) 2-Butanone-d5	4.615	46	213748	52.073	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.140%
24) Chloroform-d	5.055	84	178497	4.483	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
26) 1,2-Dichloroethane-d4	5.695	65	96902	5.267	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	5.721	84	348240	4.154	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.685	67	106033	4.197	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.000%
41) Toluene-d8	7.894	98	269277	3.638	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.800%
43) trans-1,3-Dichloroprop...	8.177	79	38697	3.994	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.800%
46) 2-Hexanone-d5	8.627	63	187213	49.166	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.340%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	83058	4.601	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	105783	4.932	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.600%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.573	96	3130	0.186	ug/L	# 1
19) 1,1-Dichloroethane	3.859	63	243672	7.096	ug/L	98
22) cis-1,2-Dichloroethene	4.660	96	4547	0.231	ug/L	84
29) 1,1,1-Trichloroethane	5.312	97	14378	0.445	ug/L	96
34) Trichloroethene	6.537	95	114318	5.226	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 17 23:23:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 17 23:14:39 2023
Response via : Initial Calibration

