

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111820\
 Data File : VU041313.D
 Acq On : 18 Nov 2020 11:23
 Operator : SY/MD
 Sample : L4485-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL-WATER-07

Quant Time: Nov 18 11:48:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM111220.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 13 13:55:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.23	114	5263	0.50	ug/L	0.00
5) Chlorobenzene-d5	9.36	117	5221	0.50	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.74	152	2544	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.61	65	3997	0.50	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.00%
4) 1,2-Dichloroethane-d4	5.68	65	3671	0.46	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.00%
7) 1,2-Dichloropropane-d6	6.66	67	3141	0.49	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	98.00%
8) Toluene-d8	7.85	98	4863	0.47	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.00%
10) 1,1,2,2-Tetrachloroethane-	10.68	84	2935	0.46	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	92.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.61	62	221	0.029 ug/L #	1
6) Trichloroethene	6.51	95	175	0.032 ug/L	94
9) 1,2-Dibromoethane	8.87	107	128	0.031 ug/L #	89
12) 1,2,3-Trichloropropane	10.75	75	123	0.032 ug/L	95
13) 1,2-Dibromo-3-chloropropan	12.91	75	22	0.027 ug/L #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111820\
Data File : VU041313.D
Acq On : 18 Nov 2020 11:23
Operator : SY/MD
Sample : L4485-14
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MDL-WATER-07

Quant Time: Nov 18 11:48:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM111220.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 13 13:55:56 2020
Response via : Initial Calibration

