


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111820\
 Data File : VU041308.D
 Acq On : 18 Nov 2020 08:36
 Operator : SY/MD
 Sample : BFB34
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB34

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM111220.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Nov 18 12:13:19 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	26696	PASS
75	95	30	80	55.8	70256	PASS
95	95	100	100	100.0	126013	PASS
96	95	5	9	6.9	8664	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	72.9	91866	PASS
175	174	5	9	8.0	7384	PASS
176	174	95	101	99.1	91082	PASS
177	176	5	9	6.4	5799	PASS

m/z	Abundance
36.00	1419.0
37.10	7040.0
38.00	6494.0
39.10	2578.0
40.10	174.0
44.00	963.0
45.00	1212.0
46.00	234.0
47.00	1625.0
48.00	934.0
49.10	6404.0
50.10	26288.0
51.10	8056.0
52.10	374.0
53.00	150.0
55.00	380.0
56.00	2017.0
57.10	3781.0
58.00	176.0
60.00	1392.0
61.10	6418.0
62.10	6976.0
63.10	5584.0
64.10	915.0
65.10	1055.0
65.90	132.0
67.00	468.0
68.10	13647.0
69.10	14083.0
70.10	900.0
72.00	864.0
73.10	6644.0
74.10	24048.0
75.10	69248.0
76.10	6216.0
77.10	844.0
78.00	615.0
79.00	5277.0
80.00	1452.0
81.00	5636.0
82.00	1194.0
83.00	181.0
86.00	150.0
87.00	5104.0
88.10	4569.0
91.00	531.0
92.10	4129.0
93.10	5996.0
94.10	15245.0
95.10	123232.0
96.10	8644.0
97.20	205.0
104.00	623.0
105.00	171.0
106.00	697.0
107.00	139.0
111.00	156.0
113.00	169.0
114.90	170.0
116.00	602.0
117.00	1043.0
118.00	602.0
119.00	923.0
128.00	576.0
129.00	246.0
130.10	435.0
131.00	245.0
135.00	293.0
137.00	247.0
140.00	128.0
141.00	1938.0
142.20	166.0
143.00	1960.0
144.10	104.0
145.00	661.0
145.90	200.0
147.90	308.0
150.00	142.0
155.10	270.0
157.10	251.0
159.10	129.0
171.10	276.0
172.10	1118.0
174.10	88464.0
175.10	7321.0
176.10	87592.0
177.10	5719.0
178.00	175.0

Instrument :
MSVOA_U
ClientSampleId :
BFB34

m/z	Abundance
36.00	1422.0
37.10	7176.0
38.10	6657.0
39.10	2614.0
40.00	160.0
44.00	992.0
45.00	1217.0
46.00	229.0
47.00	1643.0
48.00	868.0
49.10	6468.0
50.10	27648.0
51.10	8638.0
52.10	394.0
53.00	138.0
55.00	421.0
56.00	2111.0
57.10	3987.0
58.00	199.0
60.00	1436.0
61.10	6575.0
62.10	7326.0
63.10	5835.0
64.10	904.0
65.10	953.0
66.00	123.0
67.00	437.0
68.10	14286.0
69.10	15000.0
70.10	985.0
72.00	828.0
73.10	6772.0
74.10	24928.0
75.10	72808.0
76.10	6235.0
77.10	828.0
78.00	598.0
79.00	5428.0
80.00	1500.0
81.00	5908.0
82.00	1192.0
83.00	211.0
87.10	5396.0
88.10	4670.0
91.00	597.0
92.10	4286.0
93.10	6445.0
94.10	16544.0
95.10	130640.0
96.10	9034.0
97.20	243.0
104.00	731.0
105.00	203.0
106.00	676.0
107.00	151.0
109.90	110.0
111.00	146.0
113.00	172.0
114.80	203.0
116.00	620.0
117.00	1065.0
118.00	578.0
119.00	1026.0
123.90	100.0
128.00	606.0
129.00	244.0
130.00	462.0
131.00	285.0
135.00	320.0
137.00	252.0
140.10	145.0
141.00	1965.0
142.10	188.0
143.10	2004.0
144.10	103.0
145.00	642.0
145.90	189.0
148.00	355.0
150.00	152.0
155.10	285.0
157.10	268.0
159.10	109.0
171.10	271.0
172.10	1050.0
174.10	94888.0
175.10	7754.0
176.10	94520.0
177.10	5999.0
178.00	164.0

Instrument :
MSVOA_U
ClientSampleId :
BFB34

m/z	Abundance
36.00	1264.0
37.00	6565.0
38.10	6121.0
39.10	2412.0
40.00	165.0
44.00	906.0
45.00	1104.0
46.00	174.0
47.00	1450.0
48.00	783.0
49.10	5794.0
50.10	26152.0
51.10	8277.0
52.10	361.0
53.00	131.0
55.00	379.0
56.00	1998.0
57.10	3859.0
58.00	172.0
60.10	1348.0
61.00	6136.0
62.10	6798.0
63.10	5629.0
64.10	773.0
65.10	823.0
67.00	359.0
68.10	13378.0
69.10	14232.0
70.10	1047.0
72.00	762.0
73.10	6304.0
74.10	23192.0
75.10	68712.0
76.10	5704.0
77.10	762.0
78.00	503.0
79.00	5100.0
80.00	1407.0
81.00	5585.0
82.00	1111.0
83.10	208.0
87.10	5029.0
88.10	4464.0
91.00	594.0
92.10	3925.0
93.10	6078.0
94.10	16204.0
95.10	124168.0
96.10	8316.0
97.20	249.0
104.00	769.0
105.00	233.0
106.00	611.0
107.00	144.0
109.90	107.0
110.90	130.0
113.10	140.0
114.80	212.0
116.00	557.0
117.00	960.0
117.90	526.0
119.00	951.0
128.00	545.0
129.00	232.0
130.00	436.0
131.00	273.0
135.00	279.0
136.90	242.0
140.10	152.0
141.00	1777.0
142.00	220.0
143.10	1999.0
145.00	528.0
146.10	158.0
148.00	374.0
150.00	147.0
155.10	288.0
157.10	267.0
159.10	105.0
160.90	113.0
171.20	241.0
172.10	972.0
174.10	92248.0
175.10	7079.0
176.10	91136.0
177.10	5681.0
178.00	141.0

Instrument :
MSVOA_U
ClientSampleId :
BFB34

m/z	Abundance
44.00	202.0

Instrument :
MSVOA_U
ClientSampleId :
BFB34