

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111821\
 Data File : VU045860.D
 Acq On : 18 Nov 2021 11:31
 Operator : SY/MD
 Sample : M4686-11DL 10X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG1Z9DL

Quant Time: Nov 19 00:52:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 00:50:55 2021
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards							
1) 1,4-Difluorobenzene		6.253	114	91930	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.420	117	96266	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.815	152	44327	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.600	65	37207	5.108	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130		Recovery	= 102.200%		
7) Chloroethane-d5		1.916	69	28254	5.329	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130		Recovery	= 106.600%		
11) 1,1-Dichloroethene-d2		2.571	65	15391	5.038	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125		Recovery	= 100.800%		
20) 2-Butanone-d5		4.645	46	99977	51.221	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130		Recovery	= 102.440%		
24) Chloroform-d		5.066	84	59266	4.884	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125		Recovery	= 97.600%		
26) 1,2-Dichloroethane-d4		5.706	65	34549	4.852	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130		Recovery	= 97.000%		
32) Benzene-d6		5.732	84	121910	4.599	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125		Recovery	= 92.000%		
36) 1,2-Dichloropropane-d6		6.693	67	38674	4.629	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140		Recovery	= 92.600%		
41) Toluene-d8		7.902	98	102846	4.289	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130		Recovery	= 85.800%		
43) trans-1,3-Dichloroprop...		8.182	79	14874	4.385	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130		Recovery	= 87.800%		
46) 2-Hexanone-d5		8.635	63	64440	42.211	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130		Recovery	= 84.420%		
56) 1,1,2,2-Tetrachloroeth...		10.758	84	33593	4.759	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120		Recovery	= 95.200%		
66) 1,2-Dichlorobenzene-d4		12.195	152	39168	5.146	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120		Recovery	= 103.000%		
Target Compounds							Qvalue
12) 1,1-Dichloroethene		2.584	96	1094	0.192	ug/L	# 1
19) 1,1-Dichloroethane		3.870	63	2055	0.180	ug/L	98
22) cis-1,2-Dichloroethene		4.668	96	1599	0.242	ug/L	83
34) Trichloroethene		6.546	95	42945	6.203	ug/L	99
47) Tetrachloroethene		8.555	164	11162	2.303	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111821\
Data File : VU045860.D
Acq On : 18 Nov 2021 11:31
Operator : SY/MD
Sample : M4686-11DL 10X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG1Z9DL

Quant Time: Nov 19 00:52:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 00:50:55 2021
Response via : Initial Calibration

