

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111821\
 Data File : VU045870.D
 Acq On : 18 Nov 2021 15:55
 Operator : SY/MD
 Sample : M4664-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4684

Quant Time: Nov 19 00:54:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 00:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	93980	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	96435	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	46174	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	34279	4.604	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.916	69	26374	4.866	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.400%
11) 1,1-Dichloroethene-d2	2.571	65	14140	4.528	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.600%
20) 2-Butanone-d5	4.642	46	105025	52.634	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.260%
24) Chloroform-d	5.067	84	47002	3.789	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.800%
26) 1,2-Dichloroethane-d4	5.706	65	33816	4.646	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
32) Benzene-d6	5.732	84	118387	4.458	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
36) 1,2-Dichloropropane-d6	6.694	67	38563	4.607	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.200%
41) Toluene-d8	7.902	98	99637	4.147	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	8.179	79	14488	4.264	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.200%
46) 2-Hexanone-d5	8.636	63	68328	44.680	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.360%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	33602	4.752	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.000%
66) 1,2-Dichlorobenzene-d4	12.195	152	38330	4.835	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%
Target Compounds						Qvalue
5) Vinyl chloride	1.607	62	85414	10.495	ug/L	99
12) 1,1-Dichloroethene	2.584	96	2012	0.346	ug/L	# 1
17) Methyl tert-butyl Ether	3.369	73	243629	16.086	ug/L	99
18) trans-1,2-Dichloroethene	3.356	96	17933	2.876	ug/L	92
19) 1,1-Dichloroethane	3.874	63	11381	0.974	ug/L	97
22) cis-1,2-Dichloroethene	4.671	96	610787	90.334	ug/L	97
25) Chloroform	5.092	83	12477	0.984	ug/L	99
33) Benzene	5.777	78	9132	0.339	ug/L	100
34) Trichloroethene	6.546	95	571866	82.457	ug/L	98
47) Tetrachloroethene	8.555	164	5096	1.050	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111821\
Data File : VU045870.D
Acq On : 18 Nov 2021 15:55
Operator : SY/MD
Sample : M4664-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4684

Quant Time: Nov 19 00:54:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 00:50:55 2021
Response via : Initial Calibration

