

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111823\
 Data File : VU056488.D
 Acq On : 18 Nov 2023 18:25
 Operator : MD/SY
 Sample : 05476-04
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD1

Quant Time: Nov 18 23:14:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 18 00:12:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	270195	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	274499	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	118494	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	65668	3.249	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.000%
7) Chloroethane-d5	1.908	69	66631	4.300	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.000%
11) 1,1-Dichloroethene-d2	2.560	65	31146	3.492	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.800%
20) 2-Butanone-d5	4.615	46	231056	57.336	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.680%
24) Chloroform-d	5.055	84	181937	4.654	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.695	65	94628	5.239	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.800%
32) Benzene-d6	5.721	84	329091	4.024	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	6.685	67	110888	4.499	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.000%
41) Toluene-d8	7.894	98	272474	3.773	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	8.177	79	40840	4.320	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.400%
46) 2-Hexanone-d5	8.627	63	183588	49.418	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.840%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	88352	5.016	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	107183	4.896	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111823\
Data File : VU056488.D
Acq On : 18 Nov 2023 18:25
Operator : MD/SY
Sample : 05476-04
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD1

Quant Time: Nov 18 23:14:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 18 00:12:04 2023
Response via : Initial Calibration

