

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111920\
 Data File : VU041345.D
 Acq On : 19 Nov 2020 16:02
 Operator : SY/MD
 Sample : VIBLK16
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK16

Quant Time: Nov 20 06:52:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 20 05:56:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	67618	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	68619	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	28720	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	18834	3.79	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.80%
7) Chloroethane-d5	1.92	69	17662	4.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.00%
11) 1,1-Dichloroethene-d2	2.58	63	33129	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	4.66	46	62433	30.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	60.88%
24) Chloroform-d	5.08	84	49645	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
26) 1,2-Dichloroethane-d4	5.72	65	31253	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
32) Benzene-d6	5.74	84	85514	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
36) 1,2-Dichloropropane-d6	6.70	67	29850	4.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.60%
41) Toluene-d8	7.90	98	71562	4.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.20%
43) trans-1,3-Dichloropropene-	8.19	79	11906	4.34	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.80%
46) 2-Hexanone-d5	8.64	63	64196	43.80	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.60%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	26835	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	28450	5.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.60%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2081	0.377 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111920\
Data File : VU041345.D
Acq On : 19 Nov 2020 16:02
Operator : SY/MD
Sample : VIBLK16
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK16

Quant Time: Nov 20 06:52:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 20 05:56:10 2020
Response via : Initial Calibration

