

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111920\
Quantitation Report (QT Reviewed)


Data File : VU041325.D
Acq On : 18 Nov 2020 20:37
Operator : SY/MD
Sample : L4790-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4424

Manual Integrations
APPROVED

MMDadoda
11/20/2020 2:21:20 PM

Quant Time: Nov 19 01:30:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 19 00:50:13 2020
Response via : Initial Calibration



Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111920\
Quantitation Report (Qedit)




Data File : VU041325.D
Acq On : 18 Nov 2020 20:37
Operator : SY/MD
Sample : L4790-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4424

Manual Integrations
APPROVED

MMDadoda
11/20/2020 2:21:20 PM

Quant Time: Nov 19 00:52:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 19 00:50:13 2020
Response via : Initial Calibration



TIC: VU041325.D

(13) Acetone (T)

2.655min (-0.013) 0.72ug/L

response 933

Ion	Exp%	Act%
43.05	100	100
58.05	12.70	49.84#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111920\
Quantitation Report (Qedit)




Data File : VU041325.D
Acq On : 18 Nov 2020 20:37
Operator : SY/MD
Sample : L4790-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4424

Manual Integrations
APPROVED

MMDadoda
11/20/2020 2:21:20 PM

Quant Time: Nov 19 00:52:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 19 00:50:13 2020
Response via : Initial Calibration



TIC: VU041325.D

(13) Acetone (T)

2.674min (+0.006) 1.92ug/L m

response 2473

Ion	Exp%	Act%
43.05	100	100
58.05	12.70	18.80
0.00	0.00	0.00
0.00	0.00	0.00

7 MD
11/24/20

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU111920\
Quantitation Report (QT Reviewed)

Data File : VU041325.D
Acq On : 18 Nov 2020 20:37
Operator : SY/MD
Sample : L4790-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4424

Manual Integrations
APPROVED

MMDadoda
11/20/2020 2:21:20 PM

Quant Time: Nov 19 01:30:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Nov 19 00:50:13 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	64143	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	64319	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	25275	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	21036	4.46	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.20%
7) Chloroethane-d5	1.92	69	19231	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.58	63	33899	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	4.66	46	94763	48.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.42%
24) Chloroform-d	5.08	84	50252	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
26) 1,2-Dichloroethane-d4	5.72	65	31249	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
32) Benzene-d6	5.74	84	88795	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.70	67	29438	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	7.91	98	69540	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
43) trans-1,3-Dichloropropene-	8.19	79	11465	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
46) 2-Hexanone-d5	8.64	63	63359	46.12	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.24%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	26703	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	25659	5.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.00%

Target Compounds

					Qvalue
5) Vinyl chloride	1.61	62	77733	13.960	ug/L 98
13) Acetone	2.67	43	2473m	1.916	ug/L 97
17) Methyl tert-butyl Ether	3.38	73	28712	2.512	ug/L 99
18) trans-1,2-Dichloroethene	3.37	96	91149	22.414	ug/L 97
19) 1,1-Dichloroethane	3.89	63	6799	0.748	ug/L 97
22) cis-1,2-Dichloroethene	4.69	96	503379	109.604	ug/L 96
33) Benzene	5.79	78	1914	0.110	ug/L 100
34) Trichloroethene	6.55	95	3453	0.733	ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

MD
11/24/20