


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111920\
 Data File : VU041315.D
 Acq On : 18 Nov 2020 15:39
 Operator : SY/MD
 Sample : BFB14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB14

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Nov 20 05:56:10 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.3	8608	PASS
75	95	30	80	53.4	21557	PASS
95	95	100	100	100.0	40352	PASS
96	95	5	9	6.2	2482	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.7	30528	PASS
175	174	5	9	7.6	2316	PASS
176	174	95	101	96.3	29413	PASS
177	176	5	9	6.1	1804	PASS

m/z	Abundance
36.10	439.0
37.00	2162.0
38.10	2110.0
39.10	775.0
44.00	445.0
45.00	299.0
47.00	608.0
47.90	192.0
49.00	2050.0
50.10	8413.0
51.10	2533.0
51.90	135.0
55.00	131.0
56.00	650.0
57.00	1269.0
60.00	467.0
61.10	2223.0
62.10	2301.0
63.10	2054.0
64.00	391.0
65.10	363.0
67.10	164.0
68.10	4537.0
69.10	4575.0
70.00	291.0
72.00	231.0
73.10	1834.0
74.10	7063.0
75.10	21464.0
76.10	1932.0
77.10	200.0
78.10	125.0
79.00	2279.0
79.90	562.0
81.00	2086.0
81.90	486.0
87.00	1873.0
88.00	1678.0
90.90	186.0
92.10	1182.0
93.10	1696.0
94.10	5192.0
95.10	39728.0
96.10	2447.0
103.90	264.0
106.00	278.0
116.00	213.0
117.00	453.0
118.00	185.0
118.90	373.0
128.00	199.0
128.90	102.0
130.00	147.0
134.90	101.0
141.00	510.0
143.10	737.0
145.00	395.0
155.00	114.0
157.10	100.0
171.00	147.0
172.10	496.0
174.10	29616.0
175.10	2264.0
176.10	28264.0
177.10	1756.0

Instrument :
MSVOA_U
ClientSampleId :
BFB14

m/z	Abundance
36.20	452.0
37.00	2377.0
38.10	2113.0
39.10	789.0
44.10	445.0
45.00	317.0
47.00	596.0
47.90	242.0
49.00	2081.0
50.10	8914.0
51.10	2556.0
51.90	162.0
52.90	103.0
55.00	143.0
56.00	717.0
57.00	1386.0
60.00	457.0
61.10	2120.0
62.00	2319.0
63.10	2208.0
64.00	414.0
65.00	424.0
67.10	158.0
68.10	4484.0
69.10	4943.0
70.00	323.0
72.00	258.0
73.10	1893.0
74.10	7454.0
75.10	22208.0
76.10	2023.0
77.10	170.0
78.00	115.0
79.00	2223.0
80.00	623.0
81.00	2095.0
81.90	501.0
87.00	1845.0
88.00	1660.0
90.90	213.0
92.10	1168.0
93.10	1854.0
94.10	5451.0
95.10	41528.0
96.10	2559.0
103.90	285.0
105.90	290.0
116.00	237.0
117.00	501.0
117.90	188.0
118.90	459.0
128.00	183.0
130.00	156.0
140.90	575.0
143.10	786.0
145.00	400.0
155.00	119.0
171.00	188.0
172.10	525.0
174.10	31408.0
175.10	2375.0
176.10	30416.0
177.10	1851.0

Instrument :
MSVOA_U
ClientSampleId :
BFB14

m/z	Abundance
36.20	376.0
37.00	2341.0
38.00	1898.0
39.10	816.0
44.10	424.0
45.00	320.0
47.10	561.0
47.90	265.0
49.00	1798.0
50.10	8499.0
51.10	2463.0
52.00	134.0
55.00	129.0
56.00	724.0
57.00	1304.0
60.00	419.0
61.00	1844.0
62.00	2226.0
63.10	2140.0
64.00	387.0
65.00	459.0
68.10	4163.0
69.10	4736.0
70.00	296.0
72.00	257.0
73.10	1723.0
74.10	7087.0
75.10	21000.0
76.10	1947.0
76.90	156.0
79.00	1958.0
80.00	583.0
81.00	1946.0
81.90	433.0
87.00	1645.0
88.00	1520.0
91.00	211.0
92.10	1117.0
93.10	1859.0
94.10	5135.0
95.10	39800.0
96.10	2442.0
103.90	238.0
105.90	260.0
115.90	211.0
117.00	458.0
117.90	190.0
118.90	429.0
128.00	150.0
130.00	144.0
141.00	537.0
143.00	760.0
145.00	374.0
155.00	101.0
171.10	206.0
172.10	489.0
174.10	30560.0
175.10	2311.0
176.10	29560.0
177.10	1807.0

Instrument :
MSVOA_U
ClientSampleId :
BFB14

m/z	Abundance
44.00	155.0

Instrument :
MSVOA_U
ClientSampleId :
BFB14