

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111921\
 Data File : VU045886.D
 Acq On : 19 Nov 2021 11:43
 Operator : SY/MD
 Sample : M4664-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Nov 22 03:52:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 22 03:50:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	94688	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	98001	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	45788	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	30050	4.006	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.200%
7) Chloroethane-d5	1.919	69	25409	4.653	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.000%
11) 1,1-Dichloroethene-d2	2.571	65	13029	4.141	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.800%
20) 2-Butanone-d5	4.645	46	109619	54.525	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.060%
24) Chloroform-d	5.066	84	54272	4.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	5.706	65	34565	4.713	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	5.732	84	116139	4.303	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
36) 1,2-Dichloropropane-d6	6.697	67	38791	4.561	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.200%
41) Toluene-d8	7.902	98	99080	4.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
43) trans-1,3-Dichloroprop...	8.182	79	15022	4.351	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.639	63	77797	50.058	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.120%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	36124	5.027	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	12.198	152	39119	4.976	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111921\
Data File : VU045886.D
Acq On : 19 Nov 2021 11:43
Operator : SY/MD
Sample : M4664-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Nov 22 03:52:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 22 03:50:15 2021
Response via : Initial Calibration

