

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111921\
 Data File : VU045892.D
 Acq On : 19 Nov 2021 14:12
 Operator : SY/MD
 Sample : VIBLK016
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK018

Quant Time: Nov 22 03:53:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 22 03:50:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	91762	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	93823	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	44565	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	27242	3.747	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.000%
7) Chloroethane-d5	1.916	69	23293	4.401	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	2.572	65	12083	3.963	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	4.649	46	107315	55.081	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.160%
24) Chloroform-d	5.067	84	48838	4.032	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.600%
26) 1,2-Dichloroethane-d4	5.707	65	32191	4.530	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	5.732	84	106177	4.109	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.200%
36) 1,2-Dichloropropane-d6	6.694	67	35868	4.405	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.899	98	88236	3.775	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.600%
43) trans-1,3-Dichloroprop...	8.179	79	13957	4.222	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.400%
46) 2-Hexanone-d5	8.636	63	73912	49.676	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.360%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	33259	4.835	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	36232	4.735	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.800%
Target Compounds					Qvalue	
25) Chloroform	5.092	83	7339	0.593	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111921\
Data File : VU045892.D
Acq On : 19 Nov 2021 14:12
Operator : SY/MD
Sample : VIBLK016
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK018

Quant Time: Nov 22 03:53:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 22 03:50:15 2021
Response via : Initial Calibration

