

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111922\
 Data File : VU051922.D
 Acq On : 18 Nov 2022 23:56
 Operator : JC/MD
 Sample : N5684-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH1J1

Quant Time: Nov 19 00:52:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 18 22:29:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	232997	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	239056	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	91154	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	112856	4.980	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.600%
7) Chloroethane-d5	1.912	69	89391	5.078	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.600%
11) 1,1-Dichloroethene-d2	2.565	65	39828	4.903	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%
20) 2-Butanone-d5	4.616	46	200353	48.770	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.540%
24) Chloroform-d	5.060	84	165693	4.976	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
26) 1,2-Dichloroethane-d4	5.700	65	87962	5.230	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	5.726	84	349744	4.928	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
36) 1,2-Dichloropropane-d6	6.687	67	115160	5.243	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.800%
41) Toluene-d8	7.896	98	285568	4.447	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	8.179	79	36308	4.536	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.629	63	157043	46.299	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.600%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	93282	4.898	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	96413	5.524	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.400%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	2.137	101	9335	0.335	ug/L	98
17) Methyl tert-butyl Ether	3.359	73	7243	0.191	ug/L #	75
22) cis-1,2-Dichloroethene	4.661	96	11799	0.621	ug/L	93
25) Chloroform	5.086	83	13324	0.403	ug/L	97
31) Carbon tetrachloride	5.526	117	3621	0.161	ug/L	99
34) Trichloroethene	6.539	95	440350	22.693	ug/L	99
47) Tetrachloroethene	8.552	164	55701	3.745	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111922\
Data File : VU051922.D
Acq On : 18 Nov 2022 23:56
Operator : JC/MD
Sample : N5684-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH1J1

Quant Time: Nov 19 00:52:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 18 22:29:23 2022
Response via : Initial Calibration

