

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111922\
 Data File : VU051953.D
 Acq On : 19 Nov 2022 14:19
 Operator : JC/MD
 Sample : N5693-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBN91

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/21/2022
 Supervised By :Mahesh Dadoda 11/21/2022

Quant Time: Nov 19 21:44:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 19 21:39:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	224477	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	230278	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	87360	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	103212	4.727	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery =	94.600%		
7) Chloroethane-d5	1.909	69	83229	4.907	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery =	98.200%		
11) 1,1-Dichloroethene-d2	2.565	65	36402	4.651	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery =	93.000%		
20) 2-Butanone-d5	4.620	46	204413	51.647	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery =	103.300%		
24) Chloroform-d	5.060	84	152940	4.767	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	95.400%		
26) 1,2-Dichloroethane-d4	5.700	65	83258	5.138	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	102.800%		
32) Benzene-d6	5.726	84	336795	4.926	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery =	98.600%		
36) 1,2-Dichloropropane-d6	6.690	67	107825	5.096	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery =	102.000%		
41) Toluene-d8	7.896	98	277832	4.492	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery =	89.800%		
43) trans-1,3-Dichloroprop...	8.179	79	34953	4.533	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery =	90.600%		
46) 2-Hexanone-d5	8.629	63	153255	46.904	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery =	93.800%		
56) 1,1,2,2-Tetrachloroeth...	10.754	84	90667	4.943	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery =	98.800%		
66) 1,2-Dichlorobenzene-d4	12.192	152	91818	5.490	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery =	109.800%		
Target Compounds						
5) Vinyl chloride	1.604	62	21000	0.860	ug/L #	70
12) 1,1-Dichloroethene	2.578	96	11191	0.687	ug/L #	1
13) Acetone	2.620	43	265787	98.893	ug/L	100
16) Methylene chloride	3.044	84	49545	2.249	ug/L	93
18) trans-1,2-Dichloroethene	3.353	96	71537	4.098	ug/L	96
22) cis-1,2-Dichloroethene	4.661	96	11652176	636.042	ug/L	96
25) Chloroform	5.086	83	18350	0.576	ug/L	93
34) Trichloroethene	6.536	95	28813807m	1541.471	ug/L	
38) Bromodichloromethane	7.105	83	2925	0.124	ug/L #	90
47) Tetrachloroethene	8.552	164	44797	3.127	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111922\
Data File : VU051953.D
Acq On : 19 Nov 2022 14:19
Operator : JC/MD
Sample : N5693-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBN91

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/21/2022
Supervised By :Mahesh Dadoda 11/21/2022

Quant Time: Nov 19 21:44:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 19 21:39:41 2022
Response via : Initial Calibration

