

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112018\
 Data File : VU028298.D
 Acq On : 20 Nov 2018 23:19
 Operator : MD/SY
 Sample : J5986-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleID :
 H0FP8

Manual Integrations
 APPROVED

apatel
 11/21/2018 12:20:04 PM

Quant Time: Nov 21 09:03:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 21 07:07:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.60	114	120658m	5.00	ug/L	0.72
28) Chlorobenzene-d5	9.29	117	126078	5.00	ug/L	0.20
60) 1,4-Dichlorobenzene-d4	11.54	152	60340	5.00	ug/L	0.05

System Monitoring Compounds

4) Vinyl Chloride-d3	1.51	65	167	0.02	ug/L	0.12
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.40%#
7) Chloroethane-d5	1.67	69	225	0.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	0.60%#
11) 1,1-Dichloroethene-d2	2.34	63	8094	0.45	ug/L	0.06
Spiked Amount	5.000	Range	60 - 125	Recovery	=	9.00%#
20) 2-Butanone-d5	4.17	46	19452	6.40	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	12.80%#
24) Chloroform-d	4.66	84	176	0.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.20%#
26) 1,2-Dichloroethane-d4	5.39	65	233	0.02	ug/L	0.08
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.40%#
32) Benzene-d6	5.61	84	89439	2.46	ug/L	0.27
Spiked Amount	5.000	Range	70 - 125	Recovery	=	49.20%#
36) 1,2-Dichloropropane-d6	6.13	67	31372	2.38	ug/L	-0.20
Spiked Amount	5.000	Range	60 - 140	Recovery	=	47.60%#
41) Toluene-d8	7.57	98	83	0.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.00%#
43) trans-1,3-Dichloropropene-	7.67	79	36139	8.59	ug/L	-0.19
Spiked Amount	5.000	Range	55 - 130	Recovery	=	171.80%#
46) 2-Hexanone-d5	8.58	63	137520	54.47	ug/L	0.26
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.94%
57) 1,1,2,2-Tetrachloroethane-	10.52	84	49535	5.21	ug/L	0.08
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
64) 1,2-Dichlorobenzene-d4	11.90	152	48706	4.34	ug/L	0.04
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Trichlorofluoromethane	1.94	101	1624	0.11	ug/L #	66
10) 1,1,2-Trichloro-1,2,2-trif	2.34	101	6789	0.86	ug/L #	1
13) Acetone	2.34	43	10776	5.16	ug/L #	1
14) Carbon disulfide	2.34	76	11417	0.49	ug/L #	1
15) Methyl Acetate	2.69	43	16746	3.02	ug/L #	59
22) cis-1,2-Dichloroethene	4.16	96	23898	2.56	ug/L #	9
25) Chloroform	4.60	83	1233	0.06	ug/L	92
27) 1,2-Dichloroethane	5.27	62	20362	1.79	ug/L #	1
37) 1,2-Dichloropropane	6.60	63	32663	2.75	ug/L	99
49) Dibromochloromethane	8.51	129	2569	0.35	ug/L #	10
53) m,p-xylene	9.42	106	1290	0.08	ug/L	66
54) o-xylene	9.91	106	2493	0.16	ug/L	64
69) Naphthalene	13.75	128	2121	0.10	ug/L #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

