

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112018\
 Data File : VU028280.D
 Acq On : 20 Nov 2018 16:04
 Operator : MD/SY
 Sample : VU1120WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK44

Quant Time: Nov 21 07:16:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 21 07:07:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	126799	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	117364	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	48634	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	49736	5.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.20%
7) Chloroethane-d5	1.67	69	52389	6.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	127.20%
11) 1,1-Dichloroethene-d2	2.27	63	73111	3.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.00%
20) 2-Butanone-d5	4.20	46	132296	41.43	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.86%
24) Chloroform-d	4.65	84	87494	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.31	65	48966	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	5.34	84	175204	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
36) 1,2-Dichloropropane-d6	6.33	67	64600	5.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.40%
41) Toluene-d8	7.57	98	148979	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
43) trans-1,3-Dichloropropene-	7.85	79	16456	4.20	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.00%
46) 2-Hexanone-d5	8.31	63	108064	45.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.96%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	44642	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	47555	5.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU112018\
Data File : VU028280.D
Acq On : 20 Nov 2018 16:04
Operator : MD/SY
Sample : VU1120WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK44

Quant Time: Nov 21 07:16:08 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111718WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 21 07:07:52 2018
Response via : Initial Calibration

