

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U112020\
 Data File : VU041372.D
 Acq On : 20 Nov 2020 17:15
 Operator : SY/MD
 Sample : L4849-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H4407

Quant Time: Nov 21 05:35:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 21 02:42:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	58235	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	58984	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	23576	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	15824	3.69	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.80%
7) Chloroethane-d5	1.92	69	15087	4.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.40%
11) 1,1-Dichloroethene-d2	2.58	63	26829	3.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.60%
20) 2-Butanone-d5	4.64	46	75181	42.57	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.14%
24) Chloroform-d	5.08	84	43927	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.72	65	29641	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.74	84	79997	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.70	67	28339	5.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.40%
41) Toluene-d8	7.91	98	65405	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	8.19	79	10557	4.47	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.40%
46) 2-Hexanone-d5	8.64	63	50095	39.77	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.54%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	23840	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	25725	5.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.80%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.61	62	62953	12.452	ug/L	99
12) 1,1-Dichloroethene	2.59	96	1002	0.264	ug/L	# 1
13) Acetone	2.63	43	1352	1.154	ug/L	74
17) Methyl tert-butyl Ether	3.38	73	81278	7.832	ug/L	99
18) trans-1,2-Dichloroethene	3.37	96	39817	10.785	ug/L	96
19) 1,1-Dichloroethane	3.88	63	8624	1.045	ug/L	95
22) cis-1,2-Dichloroethene	4.68	96	504036	120.881	ug/L	97
25) Chloroform	5.10	83	3722	0.416	ug/L	93
33) Benzene	5.78	78	6316	0.395	ug/L	100
34) Trichloroethene	6.55	95	275725	63.824	ug/L	98
47) Tetrachloroethene	8.56	164	4715	1.477	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU112020\
Data File : VU041372.D
Acq On : 20 Nov 2020 17:15
Operator : SY/MD
Sample : L4849-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H4407

Quant Time: Nov 21 05:35:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 21 02:42:19 2020
Response via : Initial Calibration

