

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU112020\
 Data File : VU041364.D
 Acq On : 20 Nov 2020 13:59
 Operator : SY/MD
 Sample : VIBLK20
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK20

Quant Time: Nov 21 03:25:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Nov 21 02:42:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	60252	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	62419	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	24345	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	16405	3.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.00%
7) Chloroethane-d5	1.92	69	16408	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.60%
11) 1,1-Dichloroethene-d2	2.58	63	29426	3.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.20%
20) 2-Butanone-d5	4.64	46	79641	43.58	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.16%
24) Chloroform-d	5.08	84	49605	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
26) 1,2-Dichloroethane-d4	5.72	65	30961	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.74	84	84307	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	6.70	67	30366	5.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.80%
41) Toluene-d8	7.91	98	70463	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloropropene-	8.19	79	11278	4.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.40%
46) 2-Hexanone-d5	8.64	63	56037	42.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.08%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	25291	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	27215	5.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.60%


Target Compounds					Ovalue
16) Methylene chloride	3.06	84	5831	1.186 ug/L	85




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU112020\
Data File : VU041364.D
Acq On : 20 Nov 2020 13:59
Operator : SY/MD
Sample : VIBLK20
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK20

Quant Time: Nov 21 03:25:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR111620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Nov 21 02:42:19 2020
Response via : Initial Calibration





#16
Methylene chloride
Concen: 1.186 ug/L
RT: 3.06 min Scan# 534
Delta R.T. -0.00 min
Lab File: VU041364.D
Acq: 20 Nov 2020 13:59

Instrument :
MSVOA_U
ClientSampleId :
VIBLK20

Tot Ion: 84 Resp: 5831
Ion Ratio Lower Upper
84 100
86 54.7 42.8 79.6
49 119.9 99.5 184.7

