

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112023\
 Data File : VU056522.D
 Acq On : 20 Nov 2023 13:12
 Operator : MD/SY
 Sample : 05411-04RE
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0QF6RE

Quant Time: Nov 21 00:39:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 21 00:36:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	353063	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	360579	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	154378	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	46385	1.756	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	35.200%#
7) Chloroethane-d5	1.908	69	40921	2.021	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	40.400%#
11) 1,1-Dichloroethene-d2	2.563	65	23266	1.996	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	40.000%#
20) 2-Butanone-d5	4.621	46	178338	33.867	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	67.740%
24) Chloroform-d	5.059	84	136816	2.678	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	53.600%#
26) 1,2-Dichloroethane-d4	5.698	65	67445	2.857	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	57.200%#
32) Benzene-d6	5.724	84	236348	2.200	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	44.000%#
36) 1,2-Dichloropropane-d6	6.685	67	83069	2.566	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	51.400%#
41) Toluene-d8	7.894	98	184883	1.949	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	39.000%#
43) trans-1,3-Dichloroprop...	8.181	79	31373	2.526	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	50.600%#
46) 2-Hexanone-d5	8.631	63	143410	29.388	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	58.780%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	67834	2.932	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	58.600%#
66) 1,2-Dichlorobenzene-d4	12.190	152	84390	2.959	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	59.200%#
Target Compounds						
					Qvalue	
19) 1,1-Dichloroethane	3.872	63	5931	0.135	ug/L	95
29) 1,1,1-Trichloroethane	5.313	97	9562	0.231	ug/L	94
34) Trichloroethene	6.541	95	9243	0.330	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112023\
Data File : VU056522.D
Acq On : 20 Nov 2023 13:12
Operator : MD/SY
Sample : 05411-04RE
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0QF6RE

Quant Time: Nov 21 00:39:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 21 00:36:07 2023
Response via : Initial Calibration

