

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112023\
 Data File : VU056534.D
 Acq On : 20 Nov 2023 18:27
 Operator : MD/SY
 Sample : 05476-16
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAFO

Quant Time: Nov 21 00:43:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 21 00:36:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	279747	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	269428	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	122666	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	54567	2.608	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	52.200%
7) Chloroethane-d5	1.907	69	52633	3.281	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	65.600%
11) 1,1-Dichloroethene-d2	2.563	65	26505	2.870	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.400%#
20) 2-Butanone-d5	4.618	46	225688	54.092	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.180%
24) Chloroform-d	5.055	84	169516	4.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	5.695	65	84744	4.531	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
32) Benzene-d6	5.721	84	284782	3.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.000%
36) 1,2-Dichloropropane-d6	6.685	67	103226	4.267	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.894	98	222828	3.143	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	62.800%#
43) trans-1,3-Dichloroprop...	8.177	79	35530	3.829	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.600%
46) 2-Hexanone-d5	8.627	63	182312	49.998	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.000%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	85188	4.928	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	104714	4.621	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU112023\
Data File : VU056534.D
Acq On : 20 Nov 2023 18:27
Operator : MD/SY
Sample : 05476-16
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF0

Quant Time: Nov 21 00:43:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 21 00:36:07 2023
Response via : Initial Calibration

